

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.020 °C                  | 0.2 °C             | 0.220 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.436 °C                  | 20.5 °C            | 0.064 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 35 °C                      | 35.277 °C                  | 35.1 °C            | -0.177 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.366 %HR                 | 30.3 %HR           | -0.066 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 55 %HR                     | 55.345 %HR                 | 54.5 %HR           | -0.845 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 80 %HR                     | 80.570 %HR                 | 80.1 %HR           | -0.470 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |