

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.020 °C	0.2 °C	0.220 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.3 °C	-0.136 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	35.5 °C	0.223 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	30.9 %HR	0.534 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.345 %HR	55.9 %HR	0.555 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.570 %HR	78.8 %HR	-1.770 %HR	± 3.0 %HR	± 1.9E+00 %HR