

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.138 °C	10.3 °C	0.162 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.3 °C	0.224 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.7 %HR	-0.870 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.038 %HR	50.5 %HR	0.462 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.151 %HR	75.2 %HR	0.049 %HR	± 3.0 %HR	± 1.9E+00 %HR