

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-25.2 °C	-0.161 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.331 °C	0.2 °C	-0.131 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.134 °C	5.0 °C	-0.134 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.107 °C	10.0 °C	-0.107 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.4 °C	-0.164 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.401 °C	40.3 °C	-0.101 °C	± 0.5 °C	± 4.0E-01 °C

