

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-25.5 °C	-0.461 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	0.331 °C	0.5 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.134 °C	5.4 °C	0.266 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.107 °C	10.4 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.7 °C	0.351 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.9 °C	0.336 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.401 °C	40.7 °C	0.299 °C	± 0.5 °C	± 4.0E-01 °C

