

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.171 °C	0.2 °C	0.371 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.813 °C	5.2 °C	0.387 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.977 °C	10.3 °C	0.323 °C	± 1.0 °C	± 4.0E-01 °C