

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.280 °C	0.5 °C	0.220 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.3 °C	0.265 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.477 °C	35.9 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.081 %HR	33.0 %HR	1.919 %HR	± 3.0 %HR	± 1.9E+00 %HR