

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.770400	-20.130	± 1.8	± 1.5E+00 mm					
0 a 0.8	0.820100	-19.830	± 1.8	± 1.5E+00 mm					
0 a 0.9	0.920500	-20.220	± 1.9	± 1.5E+00 mm					
0 a 1	1.020700	-20.420	± 2	± 1.5E+00 mm					
0 a 1.2	1.220100	-19.810	± 2.1	± 1.5E+00 mm					
0 a 1.4	1.420300	-20.000	± 2.3	± 1.5E+00 mm					
0 a 1.6	1.620800	-20.490	± 2.4	± 1.5E+00 mm					
0 a 1.85	1.870200	-19.880	± 2.6	± 1.5E+00 mm					
0 a 2	2.020500	-20.170	± 2.8	± 1.5E+00 mm					