

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.554 °C	12.7 °C	0.146 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	26.7 °C	-0.354 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	40.090 °C	40.2 °C	0.110 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	50.601 %HR	52.6 %HR	1.999 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	70.141 %HR	72.1 %HR	1.959 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.227 %HR	94.1 %HR	1.873 %HR	± 3.0 %HR	± 1.9E+00 %HR