

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.261 °C	0.2 °C	0.461 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.013 °C	5.2 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	33.6 %HR	3.439 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.468 %HR	52.8 %HR	2.332 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.182 %HR	72.4 %HR	2.218 %HR	± 3.0 %HR	± 1.9E+00 %HR