

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.554 °C	12.6 °C	0.046 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	26.9 °C	-0.154 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	40.090 °C	39.9 °C	-0.190 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	50.601 %HR	53.5 %HR	2.899 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	70.141 %HR	72.4 %HR	2.259 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.227 %HR	93.0 %HR	0.773 %HR	± 3.0 %HR	± 1.9E+00 %HR