

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.261 °C	0.2 °C	0.461 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.013 °C	5.4 °C	0.387 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.3 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	32.0 %HR	1.839 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.468 %HR	50.7 %HR	0.232 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.182 %HR	71.2 %HR	1.018 %HR	± 3.0 %HR	± 1.9E+00 %HR