

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.554 °C	12.9 °C	0.346 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	26.8 °C	-0.254 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	40.090 °C	39.9 °C	-0.190 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	50.601 %HR	53.4 %HR	2.799 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	70.141 %HR	70.4 %HR	0.259 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.227 %HR	94.7 %HR	2.473 %HR	± 3.0 %HR	± 1.9E+00 %HR