

Certificado No. (Certificate #):		59393	Equipo (Instrument): INDICADOR DE DIAL		
Fecha de Calibración (Calibration Date):		2025-07-28	Marca (Brand): BAKER		
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
1.0000	mm	1.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
1.5000	mm	1.50 mm	0.0000 µm	± µm	± 5.8E+00 µm
2.0000	mm	2.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
4.0000	mm	4.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
5.0000	mm	5.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
7.0000	mm	7.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
8.0000	mm	8.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
9.0000	mm	9.00 mm	0.0000 µm	± µm	± 5.8E+00 µm
10.0000	mm	10.00 mm	0.0000 µm	± µm	± 5.8E+00 µm