

| Certificado No. (Certificate #):         |  | 59393      | Equipo (Instrument): INDICADOR DE DIAL |                      |            |
|------------------------------------------|--|------------|----------------------------------------|----------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-07-28 | Marca (Brand): BAKER                   |                      |            |
| Patrón (Standard)                        |  | UBP (UUT)  | Error (Error)                          | Exactitud (Accuracy) |            |
| 1.00 mm                                  |  | 1.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 1.50 mm                                  |  | 1.50 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 2.00 mm                                  |  | 2.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 4.00 mm                                  |  | 4.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 5.00 mm                                  |  | 5.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 7.00 mm                                  |  | 7.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 8.00 mm                                  |  | 8.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 9.00 mm                                  |  | 9.00 mm    | 0.00 µm                                | ±                    | 5.8E+00 µm |
| 10.00 mm                                 |  | 10.00 mm   | 0.00 µm                                | ±                    | 5.8E+00 µm |