

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
5.0068 ml	5 ml	-0.007 ml	± 0.1 ml	±	1.7E-01 µl
50.0483 ml	50 ml	-0.048 ml	± 0.1 ml	±	1.6E+00 µl
100.0608 ml	100 ml	-0.061 ml	± 0.1 ml	±	3.2E+00 µl