

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1.00 mm           | 0.98 mm   | -20.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 1.50 mm           | 1.48 mm   | -20.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 3.00 mm           | 2.98 mm   | -20.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 5.00 mm           | 4.99 mm   | -10.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 10.00 mm          | 9.97 mm   | -30.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 15.00 mm          | 14.97 mm  | -30.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 20.00 mm          | 19.97 mm  | -30.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 50.00 mm          | 49.97 mm  | -30.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |
| 100.00 mm         | 99.98 mm  | -20.00 µm     | ± 30 µm           | ±                           | 5.8E+00 µm |