

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	1.05 mm	50.00 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.52 mm	20.00 µm	± 30 µm	±	5.8E+00 µm
3.00 mm	3.06 mm	60.00 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	5.04 mm	40.00 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	10.03 mm	30.00 µm	± 30 µm	±	5.8E+00 µm
15.00 mm	15.07 mm	70.00 µm	± 30 µm	±	5.8E+00 µm
20.00 mm	19.94 mm	-60.00 µm	± 30 µm	±	5.8E+00 µm
50.00 mm	49.91 mm	-90.00 µm	± 30 µm	±	5.8E+00 µm
100.00 mm	99.95 mm	-50.00 µm	± 30 µm	±	5.8E+00 µm