

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.644 °C	12.6 °C	-0.044 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	26.954 °C	26.7 °C	-0.254 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	40.391 °C	39.7 °C	-0.691 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	49.865 %HR	51.2 %HR	1.365 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	69.773 %HR	72.5 %HR	2.677 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.022 %HR	92.9 %HR	0.878 %HR	± 3.0 %HR	± 1.9E+00 %HR