

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.484 °C	12.7 °C	0.246 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.054 °C	27.3 °C	0.266 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.085 °C	42.2 °C	0.145 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	51.572 %HR	53.0 %HR	1.388 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	70.202 %HR	71.4 %HR	1.158 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	90.673 %HR	91.2 %HR	0.557 %HR	± 3.0 %HR	± 1.9E+00 %HR