

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 12 °C                      | 12.344 °C                  | 12.7 °C            | 0.376 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 27 °C                      | 27.355 °C                  | 27.2 °C            | -0.155 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 42 °C                      | 40.792 °C                  | 40.5 °C            | -0.292 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 52 %HR                     | 50.202 %HR                 | 51.9 %HR           | 1.698 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 72 %HR                     | 70.775 %HR                 | 72.2 %HR           | 1.435 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| # 92 %HR                   | 92.636 %HR                 | 93.6 %HR           | 0.964 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |