

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 12 °C                      | 12.033 °C                  | 12.3 °C            | 0.287 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 27 °C                      | 26.954 °C                  | 27.3 °C            | 0.366 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 42 °C                      | 41.925 °C                  | 42.3 °C            | 0.375 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 52 %HR                     | 50.713 %HR                 | 52.8 %HR           | 2.087 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 72 %HR                     | 70.570 %HR                 | 72.9 %HR           | 2.350 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| # 92 %HR                   | 91.716 %HR                 | 92.6 %HR           | 0.884 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |