

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.101 °C	0.3 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.903 °C	5.3 °C	0.397 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.867 °C	10.3 °C	0.433 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.130 %HR	32.7 %HR	2.570 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	51.0 %HR	1.728 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.325 %HR	71.8 %HR	1.475 %HR	± 3.0 %HR	± 1.9E+00 %HR