

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.101 °C	0.3 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.903 °C	5.2 °C	0.297 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.867 °C	10.3 °C	0.433 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.130 %HR	33.2 %HR	3.070 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.272 %HR	50.5 %HR	1.228 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.325 %HR	73.3 %HR	2.975 %HR	± 3.0 %HR	± 1.9E+00 %HR