

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.055 °C	2.0 °C	-0.055 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.631 °C	7.5 °C	-0.131 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.756 °C	9.5 °C	-0.256 °C	± 0.5 °C	± 4.0E-01 °C