

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
# -30 °C	-30.053 °C	-30.5 °C	-0.447 °C	± 0.5 °C	± 4.0E-01 °C
# -25 °C	-25.150 °C	-25.5 °C	-0.350 °C	± 0.5 °C	± 4.0E-01 °C
# -15 °C	-15.212 °C	-15.5 °C	-0.288 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.630 °C	-11.0 °C	-0.370 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.754 °C	1.5 °C	-0.254 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.022 °C	7.5 °C	-0.522 °C	± 0.5 °C	± 4.0E-01 °C