

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.053 °C	-30.3 °C	-0.247 °C	± 0.5 °C	± 4.0E-01 °C
# -25 °C	-25.150 °C	-25.3 °C	-0.150 °C	± 0.5 °C	± 4.0E-01 °C
# -15 °C	-15.212 °C	-15.3 °C	-0.088 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.529 °C	-10.5 °C	0.029 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.764 °C	2.1 °C	0.336 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.232 °C	8.5 °C	0.268 °C	± 0.5 °C	± 4.0E-01 °C