

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.693 °C	16.0 °C	0.307 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.460 °C	25.7 °C	0.240 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.571 °C	29.6 °C	0.029 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.867 %HR	63.1 %HR	2.233 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.435 %HR	66.1 %HR	1.665 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	68.515 %HR	68.6 %HR	0.085 %HR	± 3.0 %HR	± 1.9E+00 %HR