

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.693 °C	16.2 °C	0.507 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.460 °C	25.9 °C	0.440 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.571 °C	29.5 °C	-0.071 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	60.867 %HR	60.5 %HR	-0.367 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.435 %HR	63.2 %HR	-1.235 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	68.515 %HR	65.5 %HR	-3.035 %HR	± 3.0 %HR	± 1.9E+00 %HR