

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.010 °C	0.1 °C	0.120 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.913 °C	5.3 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.827 °C	10.3 °C	0.493 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	30.8 %HR	1.957 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	52.9 %HR	2.320 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.899 %HR	73.8 %HR	1.901 %HR	± 3.0 %HR	± 1.9E+00 %HR