

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.081 °C                  | 0.2 °C             | 0.311 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 5 °C                       | 4.843 °C                   | 5.2 °C             | 0.327 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 10 °C                      | 9.877 °C                   | 10.3 °C            | 0.443 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.762 %HR                 | 32.6 %HR           | 2.838 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.110 %HR                 | 51.3 %HR           | 1.210 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 71.409 %HR                 | 74.2 %HR           | 2.791 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |