

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.101 °C	0.2 °C	0.291 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.813 °C	5.2 °C	0.417 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.017 °C	10.5 °C	0.443 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.302 %HR	30.9 %HR	1.598 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.122 %HR	53.9 %HR	2.778 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	73.7 %HR	3.048 %HR	± 3.0 %HR	± 1.9E+00 %HR