

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.121 °C	0.2 °C	0.321 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.913 °C	5.2 °C	0.297 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.827 °C	10.2 °C	0.383 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.028 %HR	32.6 %HR	2.572 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	51.9 %HR	2.649 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.347 %HR	73.2 %HR	1.853 %HR	± 3.0 %HR	± 1.9E+00 %HR