

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.161 °C	0.1 °C	0.281 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.913 °C	5.0 °C	0.127 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.917 °C	10.2 °C	0.323 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.832 %HR	31.1 %HR	2.268 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.662 %HR	52.2 %HR	1.538 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.857 %HR	72.3 %HR	1.443 %HR	± 3.0 %HR	± 1.9E+00 %HR