

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.8 °C	0.578 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.948 °C	25.7 °C	0.752 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.519 °C	51.0 °C	0.481 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	24.128 %HR	22.1 %HR	-2.028 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.282 %HR	48.2 %HR	-1.082 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.742 %HR	73.8 %HR	-0.942 %HR	± 5.0 %HR	± 1.9E+00 %HR