

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
15 °C	14.720 °C	15.0 °C	0.300 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.4 °C	0.271 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.616 °C	50.0 °C	0.414 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	23.617 %HR	21.9 %HR	-1.717 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.407 %HR	48.3 %HR	-2.107 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.640 %HR	73.4 %HR	-1.240 %HR	± 5.0 %HR	± 1.9E+00 %HR