

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.4 °C	0.309 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.948 °C	25.2 °C	0.282 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	49.115 °C	49.3 °C	0.225 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.231 %HR	22.8 %HR	-1.391 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.691 %HR	48.5 %HR	-1.241 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.333 %HR	73.1 %HR	-1.233 %HR	± 5.0 %HR	± 1.9E+00 %HR