

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.322 °C                  | 15.1 °C            | -0.222 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.450 °C                  | 25.0 °C            | -0.450 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 49.716 °C                  | 50.4 °C            | 0.684 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 25 %HR                     | 23.924 %HR                 | 21.5 %HR           | -2.424 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.100 %HR                 | 48.9 %HR           | -1.200 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 74.742 %HR                 | 73.9 %HR           | -0.842 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |