

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.5 °C	0.379 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.948 °C	24.8 °C	-0.148 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	50.8 °C	0.582 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	23.515 %HR	20.1 %HR	-3.415 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.304 %HR	47.6 %HR	-2.704 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.458 %HR	74.7 %HR	-0.758 %HR	± 5.0 %HR	± 1.9E+00 %HR