

Certificado No. (Certificate #):		59471		Equipo (Instrument): Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-04-02		Marca (Brand): Shars	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	25 mm	25.01 mm	10 µm	± 30 µm	± 8.1E+00 µm
Exterior / (Exterior)	50 mm	50.01 mm	10 µm	± 30 µm	± 8.1E+00 µm
Exterior / (Exterior)	100 mm	100.02 mm	20 µm	± 30 µm	± 8.1E+00 µm
Exterior / (Exterior)	280 mm	280.03 mm	30 µm	± 40 µm	± 8.1E+00 µm
Exterior / (Exterior)	300 mm	300.03 mm	30 µm	± 40 µm	± 8.1E+00 µm
Interior / (Interior)	50 mm	50.01 mm	10 µm	± 30 µm	± 8.2E+00 µm
Interior / (Interior)	100 mm	100.02 mm	20 µm	± 30 µm	± 8.2E+00 µm
Profundidad / (Depth)	- mm	- mm	- µm	± - µm	± - µm
Profundidad / (Depth)	- mm	- mm	- µm	± - µm	± - µm