

| Certificado No. (Certificate #):         |  | 59472      |  | Equipo (Instrument): |  | Reloj Palpador       |  |                             |            |
|------------------------------------------|--|------------|--|----------------------|--|----------------------|--|-----------------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-07-31 |  | Marca (Brand):       |  | Mitutoyo             |  |                             |            |
| Patrón (Standard)                        |  | UBP (UUT)  |  | Error (Error)        |  | Exactitud (Accuracy) |  | Incertidumbre (Uncertainty) |            |
| 0.00100 in                               |  | 0.0010 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.00200 in                               |  | 0.0020 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.00300 in                               |  | 0.0030 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.00400 in                               |  | 0.0040 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.00500 in                               |  | 0.0050 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.00600 in                               |  | 0.0060 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.01000 in                               |  | 0.0100 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.8E-05 in |
| 0.01400 in                               |  | 0.0140 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.9E-05 in |
| 0.01600 in                               |  | 0.0160 in  |  | 0.0000 in            |  |                      |  | ±                           | 5.9E-05 in |