

Certificado No. (Certificate #):		59474		Equipo (Instrument):		Micrómetro de Profundidad			
Fecha de Calibración (Calibration Date):		2025-07-04		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
0.050 in		0.050 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.100 in		0.100 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.110 in		0.110 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.120 in		0.120 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.130 in		0.130 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.150 in		0.150 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.180 in		0.180 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.200 in		0.200 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.300 in		0.300 in		0.000 in		± 0.0003 in		±	5.8E-04 in
0.500 in		0.500 in		0.000 in		± 0.0003 in		±	5.8E-04 in
1.000 in		1.000 in		0.000 in		± 0.0003 in		±	5.8E-04 in