

Certificado No. (Certificate #):		59475		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-07-04		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
1.0000 in		1.00000 in		0.00000 in		± 0.0001 in		±	2.9E-05 in
1.0500 in		1.05010 in		0.00010 in		± 0.0002 in		±	2.9E-05 in
1.1001 in		1.10015 in		0.00005 in		± 0.0002 in		±	2.9E-05 in
1.1009 in		1.10095 in		0.00005 in		± 0.0002 in		±	2.9E-05 in
1.3000 in		1.30010 in		0.00010 in		± 0.0002 in		±	2.9E-05 in
1.5000 in		1.50015 in		0.00015 in		± 0.0002 in		±	2.9E-05 in
1.6000 in		1.60010 in		0.00010 in		± 0.0002 in		±	2.9E-05 in
1.7000 in		1.70015 in		0.00015 in		± 0.0002 in		±	2.9E-05 in
1.8000 in		1.80015 in		0.00015 in		± 0.0002 in		±	2.9E-05 in
1.9000 in		1.90015 in		0.00015 in		± 0.0002 in		±	2.9E-05 in
2.0000 in		2.00010 in		0.00010 in		± 0.0002 in		±	2.9E-05 in