

Certificado No. (Certificate #):		59477		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-07-04		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
2.00000 in		2.00015 in		0.00015 in		± 0.0002 in		±	3.0E-05 in
2.05000 in		2.05015 in		0.00015 in		± 0.0002 in		±	3.0E-05 in
2.10010 in		2.10020 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
2.10090 in		2.10105 in		0.00015 in		± 0.0002 in		±	3.0E-05 in
2.30000 in		2.30005 in		0.00005 in		± 0.0002 in		±	3.0E-05 in
2.50000 in		2.50010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
2.60000 in		2.60010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
2.70000 in		2.70010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
2.80000 in		2.80010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
2.90000 in		2.90010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in
3.00000 in		3.00010 in		0.00010 in		± 0.0002 in		±	3.0E-05 in