

| Certificado No. (Certificate #):         |  | 59477      |  | Equipo (Instrument): |  | Micrómetro de Exterior |  |                             |            |
|------------------------------------------|--|------------|--|----------------------|--|------------------------|--|-----------------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-07-31 |  | Marca (Brand):       |  | Mitutoyo               |  |                             |            |
| Patrón (Standard)                        |  | UBP (UUT)  |  | Error (Error)        |  | E.M.P (Tolerance)      |  | Incertidumbre (Uncertainty) |            |
| 2.00000 in                               |  | 2.00015 in |  | 0.00015 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.05000 in                               |  | 2.05015 in |  | 0.00015 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.10010 in                               |  | 2.10020 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.10090 in                               |  | 2.10105 in |  | 0.00015 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.30000 in                               |  | 2.30005 in |  | 0.00005 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.50000 in                               |  | 2.50010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.60000 in                               |  | 2.60010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.70000 in                               |  | 2.70010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.80000 in                               |  | 2.80010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 2.90000 in                               |  | 2.90010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |
| 3.00000 in                               |  | 3.00010 in |  | 0.00010 in           |  | ± 0.0002 in            |  | ±                           | 3.0E-05 in |