

| Certificado No. (Certificate #):         |  | 59478      |  | Equipo (Instrument): |  | Micrómetro de Profundidad |  |                             |            |
|------------------------------------------|--|------------|--|----------------------|--|---------------------------|--|-----------------------------|------------|
| Fecha de Calibración (Calibration Date): |  | 2025-07-04 |  | Marca (Brand):       |  | Mitutoyo                  |  |                             |            |
| Patrón (Standard)                        |  | UBP (UUT)  |  | Error (Error)        |  | E.M.P (Tolerance)         |  | Incertidumbre (Uncertainty) |            |
| 0.05000 in                               |  | 0.05010 in |  | 0.00010 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.10010 in                               |  | 0.10015 in |  | 0.00005 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.10050 in                               |  | 0.10055 in |  | 0.00005 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.10090 in                               |  | 0.10095 in |  | 0.00005 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.30000 in                               |  | 0.30010 in |  | 0.00010 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.50000 in                               |  | 0.50015 in |  | 0.00015 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.60000 in                               |  | 0.60020 in |  | 0.00020 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.70000 in                               |  | 0.70020 in |  | 0.00020 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.80000 in                               |  | 0.80020 in |  | 0.00020 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 0.90000 in                               |  | 0.90025 in |  | 0.00025 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |
| 1.00000 in                               |  | 1.00025 in |  | 0.00025 in           |  | ± 0.0003 in               |  | ±                           | 2.9E-05 in |