

Certificado No. (Certificate #):		59478		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-07-31		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
0.05000 in		0.05010 in		0.00010 in		± 0.0001 in		±	2.9E-05 in
0.10010 in		0.10015 in		0.00005 in		± 0.0001 in		±	2.9E-05 in
0.10050 in		0.10055 in		0.00005 in		± 0.0001 in		±	2.9E-05 in
0.10090 in		0.10095 in		0.00005 in		± 0.0001 in		±	2.9E-05 in
0.30000 in		0.30010 in		0.00010 in		± 0.0001 in		±	2.9E-05 in
0.50000 in		0.50015 in		0.00015 in		± 0.0001 in		±	2.9E-05 in
0.60000 in		0.60020 in		0.00020 in		± 0.0001 in		±	2.9E-05 in
0.70000 in		0.70020 in		0.00020 in		± 0.0001 in		±	2.9E-05 in
0.80000 in		0.80020 in		0.00020 in		± 0.0001 in		±	2.9E-05 in
0.90000 in		0.90025 in		0.00025 in		± 0.0001 in		±	2.9E-05 in
1.00000 in		1.00025 in		0.00025 in		± 0.0001 in		±	2.9E-05 in