

Certificado No. (Certificate #):		59479	Equipo (Instrument): Calibrador Pie de Rey		
Fecha de Calibración (Calibration Date):		2025-07-31	Marca (Brand): Mitutoyo		
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1 in	1.00 in	0 in	± 0.0011811 in	± 3.4E-04 in
Exterior / (Exterior)	2 in	1.50 in	0 in	± 0.0011811 in	± 3.4E-04 in
Exterior / (Exterior)	3 in	3.00 in	0 in	± 0.0011811 in	± 3.4E-04 in
Exterior / (Exterior)	4 in	4.00 in	0 in	± 0.0011811 in	± 3.4E-04 in
Exterior / (Exterior)	6 in	6.00 in	0 in	± 0.0011811 in	± 3.4E-04 in
Interior / (Interior)	1 in	1.00 in	0 in	± 0.0011811 in	± 3.7E-04 in
Interior / (Interior)	4 in	4.00 in	0 in	± 0.0011811 in	± 3.7E-04 in
Profundidad / (Depth)	1 in	1.00 in	0 in	± 0.0011811 in	± 3.3E-04 in
Profundidad / (Depth)	4 in	4.00 in	0 in	± 0.0011811 in	± 3.3E-04 in