

Certificado No. (Certificate #):		59479		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-07-31		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)		
Exterior / (Exterior)	1.0000 in	1.0000 in	0.0000 in	± 0.0011811 in	±	3.4E-04	in
Exterior / (Exterior)	1.5000 in	1.5000 in	0.0000 in	± 0.0011811 in	±	3.4E-04	in
Exterior / (Exterior)	3.0000 in	3.0000 in	0.0000 in	± 0.0011811 in	±	3.4E-04	in
Exterior / (Exterior)	4.0000 in	4.0000 in	0.0000 in	± 0.0011811 in	±	3.4E-04	in
Exterior / (Exterior)	6.0000 in	6.0000 in	0.0000 in	± 0.0011811 in	±	3.4E-04	in
Interior / (Interior)	1.0000 in	0.9995 in	-0.0005 in	± 0.0011811 in	±	3.7E-04	in
Interior / (Interior)	4.0000 in	3.9995 in	-0.0005 in	± 0.0011811 in	±	3.7E-04	in
Profundidad / (Depth)	1.0000 in	0.9990 in	-0.0010 in	± 0.0011811 in	±	3.3E-04	in
Profundidad / (Depth)	4.0000 in	4.0000 in	0.0000 in	± 0.0011811 in	±	3.3E-04	in