

Certificado No. (Certificate #):		59480		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-07-31		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)		
Exterior / (Exterior)	1 in	1.00 in	0 in	± 0.0011811 in	±	3.5E-04	in
Exterior / (Exterior)	2 in	2.00 in	0 in	± 0.0011811 in	±	3.5E-04	in
Exterior / (Exterior)	4 in	4.00 in	0 in	± 0.0011811 in	±	3.5E-04	in
Exterior / (Exterior)	6 in	6.00 in	0 in	± 0.0011811 in	±	3.5E-04	in
Exterior / (Exterior)	8 in	8.00 in	0 in	± 0.0015748 in	±	3.5E-04	in
Interior / (Interior)	2 in	2.00 in	0 in	± 0.0011811 in	±	3.7E-04	in
Interior / (Interior)	6 in	6.00 in	0 in	± 0.0011811 in	±	3.7E-04	in
Profundidad / (Depth)	2 in	2.00 in	0 in	± 0.0011811 in	±	3.3E-04	in
Profundidad / (Depth)	6 in	6.00 in	0 in	± 0.0011811 in	±	3.3E-04	in